



Glideabout Comode

UTOC0001

User Manual

Glideabout Comode

Description:

Ultimate Healthcare's Glideabout Commode is a versatile product allowing the safe movement of user. To ease transfers, it features 4 braked castors and detachable arms. Fitted with upholstered seat base and backrest it provides comfort. The waste pot can be removed from commode whilst the user is seated and comes with a carry handle and lid to avoid spillage. Foot rests are height adjustable, swing away and detachable.

Assembly:

Having unpacked Glideabout Commode, insert the backrest into the frame and secure using the two hand screws provided. Foot rests are detachable and simply clip onto the brackets mounted on either side of the commode's frame.

Instructions for Use:

ATTENTION: This products should always be used under supervision. If any parts are found to be damaged do not use, please contact the supplier.

Castors –

- Each castor can be braked, using your foot push down on the castor's lever, this will prevent the wheel from rotating and swivelling.
- To release the brake the lever should be pushed upward, again using your foot.
- Brakes should always be applied with the castors facing forward when entering or leaving the chair.
- With the brakes applied the Glideabout may still slide on a smooth floor or if it is tipped forward by the user. Supervision is essential when transferring in and out of the commode.

Armrests –

- The armrests slot into brackets mounted either side of the commode.
- To remove an armrest press down on the front mounting spring clip with one hand and pull armrest upward using your other hand.
- When re-fitting an armrest, ensure that the spring clip is locked in place and armrest secure.

Footrests -

- To adjust the height of footrests, first loosen the hexagonal bolt on bottom of the hangar bracket tube until the footrest can move smoothly, slide footplate to the required height and re-tighten the hexagonal bolt until the footrest is rigid.
- For access the footrest's can be swung round to do so release black plastic catches.
- Alternatively foot rests can also be swivelled up vertically.

Waste Pot –

- To remove and empty waste pot, slide horizontally out towards the back of commode.

Maintenance:

Daily before use check the effective operation of the brakes and the secure fixing of the castors to the legs. Also check that the chair is stable, i.e. all four castors are touching the floor and none of the legs are bent. Ensure the commode arms are located correctly / secure.

Cleaning:

To clean, wipe over with a soft cloth. Use general purpose neutral detergent and warm water. **DO NOT use harsh abrasives.**

To disinfect clean as above and wipe over with solution of one thousand parts per million available chlorine disinfectant. Do not use any decontamination process that involves immersion.

Product Spec:

Maximum User Weight:190kg (30stone)

Seat Width:430mm, Seat Depth:430mm, SeatHeight:510mm

Manufacturers Contact Details:

Ultimate Healthcare Ltd, Calmore Industrial Estate, Totton, Southampton, Hampshire, SO40 3XJ

Customer Service Tel: 0870 428 6295